

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 103

Burettes, 10 ml, straight stopc., PTFE, AS, conf. cert.




Batch: 43-19

Serial:

Measuring Nr. 34558

Batch size: 47

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0080 ml	5,9994 ml	2,0037 ml
Deviation:	0,0041 ml	0,0020 ml	0,0015 ml
Precision (CV):	0,0412 %	0,0328 %	0,0750 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	15.11.2019
Balance:	Sartorius - LE244S-OCE

