

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 010 103

**Burettes, 10 ml, straight stopc., PTFE, AS, conf. cert.**




**Batch:** 19-22

**Serial:**

**Measuring Nr.** 56514

**Batch size:** 50

Volume:	10,000 ml	6,000 ml	2,000 ml
<b>Tolerance:</b>	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
<b>Average value:</b>	10,0025 ml	5,9973 ml	2,0010 ml
<b>Deviation:</b>	0,0066 ml	0,0040 ml	0,0021 ml
<b>Precision (CV):</b>	0,0664 %	0,0658 %	0,1037 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 7

**Date:** 09.06.2022

**Balance:** Sartorius -  
LE244S-OCE

DE-M