

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 252

Burettes, 25 ml, straight stopcock, AS, conf. cert.




Batch: 10-17

Serial:

Measuring Nr.: 15144

Batch size: 25

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0030 ml	14,9965 ml	4,9887 ml
Deviation:	0,0108 ml	0,0071 ml	0,0055 ml
Precision (CV):	0,0433 %	0,0473 %	0,1102 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	19.10.2017
Balance:	Sartorius - LE244S-OCE

