

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 253 B

Burettes, 25 ml, str. stopc., amber grad., AS, conf. cert.




Batch: 18-24

Serial:

Measuring Nr. 67411

Batch size: 240

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0033 ml	14,9996 ml	5,0113 ml
Deviation:	0,0094 ml	0,0052 ml	0,0047 ml
Precision (CV):	0,0378 %	0,0347 %	0,0932 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.06.2024
Balance:	Sartorius - LE244S-OCE

