

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 010 253

**Burettes, 25 ml, str. stopcock, PTFE, AS, conf. cert.**




**Batch:** 06-21

**Serial:**

**Measuring Nr.:** 44905

**Batch size:** 100

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0036 ml	14,9977 ml	5,0028 ml
Deviation:	0,0092 ml	0,0096 ml	0,0059 ml
Precision (CV):	0,0369 %	0,0641 %	0,1178 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	05.03.2021
Balance:	Sartorius - LE244S-OCE

DE-M