

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 503 B

Burettes, 50 ml, str. stopc., amber grad., AS, conf. cert.




Batch: 38-17

Serial:

Measuring Nr. 15794

Batch size: 1100

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9947 ml	29,9734 ml	9,9916 ml
Deviation:	0,0108 ml	0,0090 ml	0,0082 ml
Precision (CV):	0,0215 %	0,0298 %	0,0820 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	23.11.2017
Balance:	Sartorius - LE244S-OCE

