




Certificate of performance

Article number 3 010 504
 Batch 03-17
 Serial number
 Measuringnr./Batch size 10106 / 10
 Balance Sartorius - LE244S-0CE
 Operator 9
 Date 14.02.2017

Burettes 50 ml class AS lateral stopcock glass-plug KB

Volume	50,0000ml	30,0000ml	10,0000ml
Tolerance	0,0500ml	0,0500ml	0,0500ml
Average volume	50,0280 ml	29,9935 ml	9,9934 ml
Deviation	0,0048 ml	0,0073 ml	0,0108 ml
Number of measurements	3	3	3



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
 97877 Wertheim
 Germany
 +49 (0) 9342 / 9301-0
 | info@witeg.de

