

Certificate of performance



Article: 3 012 053

Burettes, 5 ml, str. stopc., PTFE- plug, AS, conf. cert.




Batch: 24-16

Serial:

Measuring Nr. 23998

Batch size: 27

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	5,0026 ml	2,9985 ml	0,9995 ml
Deviation:	0,0027 ml	0,0017 ml	0,0021 ml
Precision (CV):	0,0531 %	0,0566 %	0,2123 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	12.09.2018
Balance:	Sartorius - LE244S-OCE

