

# Certificate of performance



**Article:** 3 012 053

**Burettes, 5 ml, str. stopc., PTFE- plug, AS, conf. cert.**




**Batch:** 35-21

**Serial:**

**Measuring Nr.** 49939

**Batch size:** 30

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	5,0012 ml	3,0018 ml	0,9999 ml
Deviation:	0,0021 ml	0,0017 ml	0,0010 ml
Precision (CV):	0,0416 %	0,0551 %	0,0992 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	20.10.2021
Balance:	Sartorius - LE244S-OCE

