

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 103

Burettes, 10 ml, str. stopc., PTFE-plug, AS, conf. cert.




Batch: 24-20

Serial:

Measuring Nr. 38971

Batch size: 66

Volume:	10,000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0025 ml	5,9952 ml	2,0037 ml
Deviation:	0,0071 ml	0,0049 ml	0,0030 ml
Precision (CV):	0,0713 %	0,0816 %	0,1487 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	03.07.2020
Balance:	Sartorius - LE244S-OCE

