

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 253

Burettes 25 ml, str. stopc. PTFE plug, AS, conf. cert.




Batch: 10-17

Serial:

Measuring Nr. 12052

Batch size: 100

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0107 ml	15,0050 ml	5,0018 ml
Deviation:	0,0054 ml	0,0078 ml	0,0052 ml
Precision (CV):	0,0217 %	0,0522 %	0,1044 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	12.07.2017
Balance:	Sartorius - LE244S-OCE

