

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 253

Burettes, 25 ml, str. stopc., PTFE plug, AS, conf. cert.




Batch: 40-18

Serial:

Measuring Nr. 25310

Batch size: 185

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	24,9986 ml	14,9974 ml	5,0056 ml
Deviation:	0,0101 ml	0,0059 ml	0,0045 ml
Precision (CV):	0,0405 %	0,0391 %	0,0890 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

31.10.2018

Balance:

Sartorius -
LE244S-OCE

DE-M