

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 502

Burettes, 50 ml, straight stopc., AS, conf. cert




Batch: 49-16

Serial:

Measuring Nr. 12542

Batch size: 10

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0000 ml	30,0010 ml	10,0076 ml
Deviation:	0,0097 ml	0,0086 ml	0,0089 ml
Precision (CV):	0,0194 %	0,0286 %	0,0886 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	02.08.2017
Balance:	Sartorius - LE244S-OCE

