

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 502

Burettes, 50 ml, straight stopc., AS, conf. cert




Batch: 30-17

Serial:

Measuring Nr. 15355

Batch size: 25

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9878 ml	29,9958 ml	9,9978 ml
Deviation:	0,0130 ml	0,0170 ml	0,0084 ml
Precision (CV):	0,0261 %	0,0568 %	0,0840 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	24.10.2017
Balance:	Sartorius - LE244S-OCE

