

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 012 503

**Burettes, 50 ml, str. stopc., PTFE plug, AS, conf. cert.**




**Batch:** 43-20

**Serial:**

**Measuring Nr.:** 42704

**Batch size:** 200

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
<b>Tolerance:</b>	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
<b>Average value:</b>	49,9953 ml	30,0052 ml	10,0138 ml
<b>Deviation:</b>	0,0172 ml	0,0098 ml	0,0090 ml
<b>Precision (CV):</b>	0,0344 %	0,0327 %	0,0900 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	23.11.2020
<b>Balance:</b>	Sartorius - LE244S-OCE

