

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 014 252

Burettes, 25 ml, str. stopc., amber gl., AS, conf. cert.




Batch: 32-15

Serial:

Measuring Nr. 21467

Batch size: 20

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0172 ml	15,0122 ml	5,0049 ml
Deviation:	0,0022 ml	0,0058 ml	0,0072 ml
Precision (CV):	0,0089 %	0,0386 %	0,1434 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	07.06.2018
Balance:	Sartorius - LE244S-OCE

DE-M