

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 014 503

Burettes, 50 ml, str. stopc., amber glass, AS, conf. cert.




Batch: 02-24

Serial:

Measuring Nr. 69533

Batch size: 50

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0245 ml	30,0076 ml	10,0020 ml
Deviation:	0,0090 ml	0,0075 ml	0,0036 ml
Precision (CV):	0,0180 %	0,0248 %	0,0363 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	16.10.2024
Balance:	Sartorius - MSE224S-0CE -DU

