

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 020 507

**Burettes, 50 ml, straight needle valve, AS, conf. cert.**




**Batch:** 38-17

**Serial:**

**Measuring Nr.** 50214

**Batch size:** 12

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0049 ml	29,9791 ml	10,0036 ml
Deviation:	0,0020 ml	0,0073 ml	0,0023 ml
Precision (CV):	0,0041 %	0,0243 %	0,0228 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	08.11.2021
Balance:	Sartorius - LE244S-OCE

DE-M