

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 022 507

Burettes, 50 ml, straight needle valve, AS, conf. cert.




Batch: 50-17

Serial:

Measuring Nr. 18864

Batch size: 70

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9891 ml	30,0122 ml	10,0075 ml
Deviation:	0,0105 ml	0,0098 ml	0,0091 ml
Precision (CV):	0,0210 %	0,0328 %	0,0912 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	15.03.2018
Balance:	Sartorius - LE244S-OCE

DE-M