

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 022 507

Burettes, 50 ml, straight needle valve, AS, conf. cert.




Batch: 32-18

Serial:

Measuring Nr.: 40002

Batch size: 1

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Measured value:	50,0334 ml	30,0425 ml	10,0172 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	09.09.2020
Balance:	Sartorius - LE244S-OCE

