

Certificate of performance



Article: 3 022 907

Burettes, 100 ml, straight needle valve, AS, conf. cert.




Batch: 41-17

Serial:

Measuring Nr.: 19740

Batch size: 1

Volume:	100,000 ml	60,000 ml	20,000 ml
Tolerance:	$\pm 0,1000$ ml	$\pm 0,1000$ ml	$\pm 0,1000$ ml
Measured value:	99,9617 ml	59,9861 ml	19,9959 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	10.04.2018
Balance:	Sartorius - LE244S-OCE

