

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 024 507

Burette, 50 ml, str. needle valve, amber, AS, conf. cert.




Batch: 46-20

Serial:

Measuring Nr.: 68944

Batch size: 10

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0121 ml	29,9707 ml	10,0009 ml
Deviation:	0,0096 ml	0,0062 ml	0,0048 ml
Precision (CV):	0,0192 %	0,0207 %	0,0483 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	09.09.2024
Balance:	Sartorius - LE244S-OCE

DE-M