

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 122 105

**Pellet-burettes, 10 ml, lat. needle valve, AS, conf. cert.**




**Batch:** 13-18

**Serial:**

**Measuring Nr.** 29268

**Batch size:** 4

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	9,9999 ml	5,9937 ml	1,9960 ml
Deviation:	0,0028 ml	0,0059 ml	0,0054 ml
Precision (CV):	0,0282 %	0,0991 %	0,2718 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	4
Date:	16.04.2019
Balance:	Sartorius - LE244S-OCE

