

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 254

Pettet-burettes, 25 ml, lateral Stopc., AS, conf. cert.




Batch: 24-18

Serial:

Measuring Nr. 25841

Batch size: 10

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0171 ml	15,0065 ml	5,0137 ml
Deviation:	0,0054 ml	0,0112 ml	0,0056 ml
Precision (CV):	0,0216 %	0,0744 %	0,1125 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	22.11.2018
Balance:	Sartorius - LE244S-OCE

DE-M