

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 255

Pellet-burettes, 25 ml, lat. needle valve, AS, conf. cert.




Batch: 24-18

Serial:

Measuring Nr. 26401

Batch size: 10

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0109 ml	15,0068 ml	5,0216 ml
Deviation:	0,0077 ml	0,0079 ml	0,0038 ml
Precision (CV):	0,0307 %	0,0529 %	0,0767 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	11.12.2018
Balance:	Sartorius - LE244S-OCE

