

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 258

Pellet-burettes, 25 ml, lat. needle valve, AS, conf. cert.




Batch: 11-22

Serial:

Measuring Nr. 54584

Batch size: 20

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	24,9940 ml	15,0088 ml	5,0108 ml
Deviation:	0,0102 ml	0,0064 ml	0,0044 ml
Precision (CV):	0,0407 %	0,0426 %	0,0887 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	27.04.2022
Balance:	Sartorius - LE244S-OCE

