

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 122 504

**Pellet-burettes, 50 ml, lateral stopc., AS, conf. cert.**




**Batch:** 13-19

**Serial:**

**Measuring Nr.** 34979

**Batch size:** 5

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0096 ml	29,9954 ml	10,0155 ml
Deviation:	0,0137 ml	0,0187 ml	0,0157 ml
Precision (CV):	0,0275 %	0,0625 %	0,1569 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	9
Date:	05.12.2019
Balance:	Sartorius - LE244S-OCE

DE-M