

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 508

Pellet-burettes, 50 ml, lat. needle valve, AS, conf. cert.




Batch: 01-18

Serial:

Measuring Nr. 36750

Batch size: 27

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9887 ml	30,0151 ml	10,0178 ml
Deviation:	0,0151 ml	0,0154 ml	0,0134 ml
Precision (CV):	0,0303 %	0,0514 %	0,1339 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	24.03.2020
Balance:	Sartorius - MSE224S-0CE -DU

