

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 104

Pellet-burette, 10 ml, amber glass, AS, conf. cert.




Batch: 29-20

Serial:

Measuring Nr.: 47277

Batch size: 10

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0114 ml	6,0037 ml	2,0059 ml
Deviation:	0,0040 ml	0,0073 ml	0,0046 ml
Precision (CV):	0,0398 %	0,1219 %	0,2315 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	10.06.2021
Balance:	Sartorius - MSE224S-0CE -DU

