

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 124 105

**Pellet-burettes, 10 ml, amber glass, AS, conf. cert.**




**Batch:** 47-15

**Serial:**

**Measuring Nr.** 17388

**Batch size:** 10

Volume:	10,000 ml	6,000 ml	2,000 ml
<b>Tolerance:</b>	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
<b>Average value:</b>	9,9946 ml	6,0020 ml	2,0068 ml
<b>Deviation:</b>	0,0046 ml	0,0044 ml	0,0025 ml
<b>Precision (CV):</b>	0,0465 %	0,0739 %	0,1260 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	9
<b>Date:</b>	23.01.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

