

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 258

Pellet-burettes, 25 ml, amber glass, AS, conf. cert.




Batch: 43-20

Serial:

Measuring Nr.: 44352

Batch size: 25

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0027 ml	14,9865 ml	4,9817 ml
Deviation:	0,0130 ml	0,0069 ml	0,0043 ml
Precision (CV):	0,0521 %	0,0460 %	0,0856 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	11.02.2021
Balance:	Sartorius - MSE224S-0CE -DU

