

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 508

Pellet-burettes, 50 ml, amber glass, AS, conf. cert.




Batch: 32-18

Serial:

Measuring Nr.: 26885

Batch size: 10

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9797 ml	29,9708 ml	10,0073 ml
Deviation:	0,0068 ml	0,0044 ml	0,0010 ml
Precision (CV):	0,0135 %	0,0147 %	0,0102 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	17.01.2019
Balance:	Sartorius - LE244S-OCE

