

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 152 508 FCH

**Universal autom. burette, 50 ml, FCH, AS, conf. cert.**

**Batch:** 30-17




**Serial:**

**Measuring Nr.** 28038

**Batch size:**

1

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Measured value:	49,9988 ml	30,0391 ml	10,0241 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

9

**Date:**

27.02.2019

**Balance:**

Sartorius -  
LE244S-OCE

DE-M