

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 372 023

Micro burettes, 2 ml, str. stopc., PTFE, AS, conf. cert.




Batch: 09-16

Serial:

Measuring Nr. 15628

Batch size: 4

Volume:	2,0000 ml	1,2000 ml	0,4000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	2,0030 ml	1,1991 ml	0,3998 ml
Deviation:	0,0015 ml	0,0025 ml	0,0017 ml
Precision (CV):	0,0731 %	0,2057 %	0,4298 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	09.11.2017
Balance:	Sartorius - LE244S-OCE

