

Certificate of performance



Article: 3 372 052

Micro burettes, 5 ml, straight stopc., AS, conf. cert.




Batch: 25-21

Serial:

Measuring Nr.: 49532

Batch size: 4

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	5,0016 ml	3,0005 ml	1,0021 ml
Deviation:	0,0029 ml	0,0019 ml	0,0016 ml
Precision (CV):	0,0578 %	0,0623 %	0,1648 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	06.10.2021
Balance:	Sartorius - LE244S-OCE

