

Certificate of performance



Article: 3 715 000

Graduated cylinder, 5:0,1 ml, A, conf. cert.




Batch: 04-21

Serial:

Measuring Nr.: 44353

Batch size: 500

Volume:	1,0000 ml	3,0000 ml	5,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	0,9794 ml	3,0024 ml	5,0065 ml
Deviation:	0,0099 ml	0,0054 ml	0,0075 ml
Precision (CV):	0,9917 %	0,1790 %	0,1498 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.02.2021
Balance:	Sartorius - LE244S-OCE

