

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 001

**Grad. cylinder, 10:0,2 ml, A, conf.cert.**




**Batch:** 11-23

**Serial:**

**Measuring Nr.** 62044

**Batch size:** 500

Volume:	2,0000 ml	6,0000 ml	10,0000 ml
<b>Tolerance:</b>	± 0,1000 ml	± 0,1000 ml	± 0,1000 ml
<b>Average value:</b>	1,9926 ml	5,9900 ml	9,9807 ml
<b>Deviation:</b>	0,0195 ml	0,0169 ml	0,0153 ml
<b>Precision (CV):</b>	0,9760 %	0,2822 %	0,1525 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	22.03.2023
<b>Balance:</b>	Sartorius - LE244S-OCE

