

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 003

**Graduated cylinder, 25:0,5 ml, A, conf. cert**




**Batch:** 16-24

**Serial:**

**Measuring Nr.** 66738

**Batch size:** 850

Volume:	5,0000 ml	15,0000 ml	25,0000 ml
Tolerance:	± 0,2500 ml	± 0,2500 ml	± 0,2500 ml
Average value:	5,0251 ml	15,0140 ml	24,9935 ml
Deviation:	0,0536 ml	0,0441 ml	0,0358 ml
Precision (CV):	1,0713 %	0,2938 %	0,1431 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	12.04.2024
Balance:	Sartorius - LE244S-OCE

