

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 005 BL

**Graduated cylinders, 50 ml, blue grad., A, conf. cert.**




**Batch:** 04-18

**Serial:**

**Measuring Nr.** 17310

**Batch size:** 280

Volume:	10,000 ml	30,000 ml	50,000 ml
<b>Tolerance:</b>	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
<b>Average value:</b>	10,0704 ml	30,0530 ml	50,0422 ml
<b>Deviation:</b>	0,0509 ml	0,0576 ml	0,0562 ml
<b>Precision (CV):</b>	0,5091 %	0,1921 %	0,1123 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	9
<b>Date:</b>	22.01.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

