

Certificate of performance



Article: 3 715 005 BL

Graduated cylinders, 50:1 ml, blue grad., A, conf. cert.




Batch: 27-19

Serial:

Measuring Nr.: 31185

Batch size: 500

Volume:	10,000 ml	30,000 ml	50,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	10,0400 ml	30,0532 ml	50,0658 ml
Deviation:	0,0608 ml	0,0501 ml	0,0519 ml
Precision (CV):	0,6082 %	0,1669 %	0,1039 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	02.07.2019
Balance:	Sartorius - LE244S-OCE

