

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 005 BL

**Graduated cylinders, 50:1 ml, blue grad., A, conf. cert.**




**Batch:** 23-23

**Serial:**

**Measuring Nr.** 63342

**Batch size:** 500

Volume:	10,000 ml	30,000 ml	50,000 ml
<b>Tolerance:</b>	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
<b>Average value:</b>	9,9764 ml	30,0069 ml	50,0312 ml
<b>Deviation:</b>	0,0513 ml	0,0493 ml	0,0462 ml
<b>Precision (CV):</b>	0,5128 %	0,1644 %	0,0923 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	13.06.2023
<b>Balance:</b>	Sartorius - LE244S-0CE

