

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 005

Graduated cylinders, 50:1 ml, A, conf. cert.




Batch: 16-18

Serial:

Measuring Nr. 20099

Batch size: 500

Volume:	10,000 ml	30,000 ml	50,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	9,9525 ml	30,0154 ml	50,0585 ml
Deviation:	0,1204 ml	0,1088 ml	0,0966 ml
Precision (CV):	1,2044 %	0,3627 %	0,1932 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	44
Date:	23.04.2018
Balance:	Sartorius - LE244S-OCE

