

Certificate of performance



Article: 3 715 005

Graduated cylinders, 50:1 ml, A, conf. cert.




Batch: 28-21

Serial:

Measuring Nr.: 47979

Batch size: 1200

Volume:	10,000 ml	30,000 ml	50,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	9,9817 ml	30,0419 ml	50,0921 ml
Deviation:	0,0794 ml	0,0599 ml	0,0370 ml
Precision (CV):	0,7941 %	0,1998 %	0,0739 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	20.07.2021
Balance:	Sartorius - LE244S-OCE

