

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 010 BL

Grad. cylinder, 100:1,0 ml, blue grad., A, conf. cert.




Batch: 10-18

Serial:

Measuring Nr. 19064

Batch size: 280

Volume:	20,000 ml	60,000 ml	100,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	19,9898 ml	60,0105 ml	99,9845 ml
Deviation:	0,0560 ml	0,0370 ml	0,0417 ml
Precision (CV):	0,2798 %	0,0616 %	0,0417 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	44
Date:	22.03.2018
Balance:	Sartorius - LE244S-OCE

