

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 010

Graduated cylinder, 100:1,0 ml, A, conf. cert.




Batch: 46-18

Serial:

Measuring Nr. 25745

Batch size: 3076

Volume:	20,000 ml	60,000 ml	100,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	19,8633 ml	59,9693 ml	100,0558 ml
Deviation:	0,1390 ml	0,0972 ml	0,0831 ml
Precision (CV):	0,6952 %	0,1620 %	0,0831 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	22.11.2018
Balance:	Sartorius - LA 1200 S

