

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 025 BL

Grad. cylinder, 250 ml, blue grad., A, conf. cert.




Batch: 14-18

Serial:

Measuring Nr. 19419

Batch size: 280

Volume:	50,0000 ml	150,0000 ml	250,0000 ml
Tolerance:	± 1,0000 ml	± 1,0000 ml	± 1,0000 ml
Average value:	49,8565 ml	150,0679 ml	250,2959 ml
Deviation:	0,1800 ml	0,0917 ml	0,0776 ml
Precision (CV):	0,3600 %	0,0611 %	0,0310 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	04.04.2018
Balance:	Sartorius - LA 1200 S

