

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 025

Graduated cylinder, 250:2 ml, A, conf. cert.




Batch: 27-22

Serial:

Measuring Nr. 57282

Batch size: 1600

Volume:	50,0000 ml	150,0000 ml	250,0000 ml
Tolerance:	± 1,0000 ml	± 1,0000 ml	± 1,0000 ml
Average value:	50,1620 ml	150,2018 ml	250,1477 ml
Deviation:	0,2012 ml	0,2047 ml	0,2516 ml
Precision (CV):	0,4023 %	0,1365 %	0,1007 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.07.2022
Balance:	Sartorius - LA 1200 S

