

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 050 BL

Grad. cylinders, 500:5 ml, blue grad., A, conf. cert.




Batch: 11-20

Serial:

Measuring Nr. 36666

Batch size: 270

Volume:	100,000 ml	300,000 ml	500,000 ml
Tolerance:	± 2,5000 ml	± 2,5000 ml	± 2,5000 ml
Average value:	100,1091 ml	299,9681 ml	499,7514 ml
Deviation:	0,4632 ml	0,3554 ml	0,3635 ml
Precision (CV):	0,4632 %	0,1185 %	0,0727 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	16.03.2020
Balance:	Sartorius - LA 1200 S

