

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 100

Graduated cylinder, 1000:10 ml, A, conf. cert.




Batch: 26-20

Serial:

Measuring Nr. 39000

Batch size: 500

Volume:	200,000 ml	600,000 ml	1000,000 ml
Tolerance:	± 5,0000 ml	± 5,0000 ml	± 5,0000 ml
Average value:	200,1780 ml	599,9674 ml	999,9877 ml
Deviation:	0,5535 ml	0,7123 ml	0,6516 ml
Precision (CV):	0,2768 %	0,1187 %	0,0652 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	06.07.2020
Balance:	Sartorius - LC 6200 S

