

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 200

**Graduated cylinder, 2000:20 ml, A, conf. cert.**




**Batch:** 32-24

**Serial:**

**Measuring Nr.** 68609

**Batch size:** 90

Volume:	400,000 ml	1200,000 ml	2000,000 ml
Tolerance:	± 10,000 ml	± 10,000 ml	± 10,000 ml
Average value:	398,5809 ml	1199,8635 ml	2001,8631 ml
Deviation:	1,4395 ml	1,2105 ml	1,2019 ml
Precision (CV):	0,3599 %	0,1009 %	0,0601 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	4
Date:	14.08.2024
Balance:	Sartorius - MSE14202S- OCE-D0

